



## King's Research Portal

DOI:

[10.1111/apa.15089](https://doi.org/10.1111/apa.15089)

*Document Version*

Peer reviewed version

[Link to publication record in King's Research Portal](#)

*Citation for published version (APA):*

Sturrock, S., Williams, E., Dassios, T., & Greenough, A. (2019). Closed loop automated oxygen control in neonates – a review: Neonatal closed loop automated oxygen control. *Acta Paediatrica*.  
<https://doi.org/10.1111/apa.15089>

### **Citing this paper**

Please note that where the full-text provided on King's Research Portal is the Author Accepted Manuscript or Post-Print version this may differ from the final Published version. If citing, it is advised that you check and use the publisher's definitive version for pagination, volume/issue, and date of publication details. And where the final published version is provided on the Research Portal, if citing you are again advised to check the publisher's website for any subsequent corrections.

### **General rights**

Copyright and moral rights for the publications made accessible in the Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognize and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the Research Portal

### **Take down policy**

If you believe that this document breaches copyright please contact [librarypure@kcl.ac.uk](mailto:librarypure@kcl.ac.uk) providing details, and we will remove access to the work immediately and investigate your claim.

**Title: Closed loop automated oxygen control in neonates – a review**

**Sarah Sturrock<sup>1</sup>, Emma Williams<sup>1</sup>, Theodore Dassios<sup>1,2</sup>, Anne Greenough<sup>1,3,4</sup>**

<sup>1</sup> Women and Children's Health, School of Life Course Sciences, Faculty of Life Sciences and Medicine, King's College London, UK; <sup>2</sup>Neonatal Intensive Care Centre, King's College Hospital, London, UK; <sup>3</sup> The Asthma UK Centre in Allergic Mechanisms of Asthma, King's College London, UK; <sup>4</sup> National Institute for Health Research (NIHR) Biomedical Research Centre based at Guy's and St Thomas' NHS Foundation Trust and King's College London, UK

**Short title:** Neonatal closed loop automated oxygen control

**Corresponding author:** Professor Anne Greenough, NICU, 4th Floor Golden Jubilee Wing, King's College Hospital, Denmark Hill, London, SE5 9RS, Tel: 0203 299 4644; fax: 0203 299 8284; email: [anne.greenough@kcl.ac.uk](mailto:anne.greenough@kcl.ac.uk)

## **ABSTRACT**

**Aims:** Neonates frequently require supplementary oxygen but may develop complications if the oxygen saturation is outside the target range. This review aimed to determine whether the algorithms used in closed loop automated oxygen control systems influenced their efficacy and whether use of the systems reduced relevant, long-term neonatal complications.

**Methods:** A literature search was conducted using PubMed and Google Scholar. The search terms were “closed loop” or “automat\*”, “oxygen” and “neonat\*”.

**Results:** Eighteen studies were identified: sixteen comparison clinical studies, an observational study and an animal study. Overall, closed loop automated oxygen control was associated with an increased percentage of time spent within the target oxygen saturation range and there were fewer manual adjustments to the inspired oxygen concentration when compared with manual oxygen control. The systems were effective in infants on non-invasive respiratory support or mechanically ventilated, but no study included term-born infants. No long-term data were available to determine if complications of oxygen toxicity were reduced.

**Conclusion:** Closed loop automated oxygen control has been shown in short term trials including preterm and low birth weight infants to improve target saturation achievement. Whether long term outcomes will be improved with their use requires investigation.

**Key words:** Closed loop automated oxygen control; neonate; bronchopulmonary dysplasia; retinopathy of prematurity

## **KEY NOTES**

Closed loop automated oxygen control systems automate the adjustment of the inspired oxygen concentrations according to peripheral oxygen saturation levels. The systems result in the delivery of supplementary oxygen more often in the desired oxygen saturation target levels with less manual intervention. As yet, however, there is no evidence that their use reduces long term complications related to supplementary oxygen, this requires investigation.

## INTRODUCTION

Neonates with respiratory distress frequently require supplementary oxygen (1), but its use can result in development of reactive oxide species (ROS) and complications such as bronchopulmonary dysplasia and retinopathy of prematurity (2). Targeting oxygen therapy to maintain oxygen saturations ( $\text{SpO}_2$ ) within a predefined range ( $\text{SpO}_2$  of 90-95%) can maximise the benefits of increased oxygen delivery to tissues whilst minimising the risk of complications (3, 4). As a consequence, in clinical practice, peripheral oxygen saturations are continuously monitored and used to guide adjustments to the inspired oxygen ( $\text{FiO}_2$ ) which are made manually by neonatal practitioners. Neonatal patients are prone to frequent fluctuations in oxygen saturations, with as many as 600 intermittent hypoxic episodes documented in one week in one study (5). Compliance with  $\text{SpO}_2$  target ranges has been shown to be variable even within the same patient over time, as well as between patients and centres (6). One quality improvement group found their target saturation range achievement was as low as 20% (7). Furthermore, narrowing the target range decreases compliance (6) and target achievement decreases as the number of patients per nurse increases (8). The shortfalls in target saturation range achievement have been suggested to be due to the clinical staff being more tolerant of  $\text{SpO}_2$  levels which are too high rather than too low (9). In one study, upper alarm limits were set too high up to 76.5% of the time, whereas lower alarm limits were set correctly 91.1% of the time (9).

Closed loop automated oxygen control systems use  $\text{SpO}_2$  values monitored in real time, to calculate and make an adjustment to the  $\text{FiO}_2$  without any human

intervention. The resultant change in  $\text{SpO}_2$  is monitored and further alterations to the  $\text{FiO}_2$  made as needed. Closed loop automated oxygen control systems may, therefore, provide a solution for the low compliance to target oxygen saturation level, reduce the need for manual adjustments (and hence workload) and decrease complications.

Closed loop automated oxygen control systems continuously monitor  $\text{SpO}_2$  and the data are fed into an algorithm which determines and executes an appropriate adjustment to the  $\text{FiO}_2$ . The result of this adjustment is monitored and further changes made if needed. The relationship between  $\text{FiO}_2$  and  $\text{SpO}_2$  in neonates needing respiratory support and supplemental oxygen is non-linear and complex (10), hence algorithms are used which reflect this. Several types of algorithm have been used.

A systematic review of clinical trials comparing closed loop automated oxygen control to manual control confirmed that these systems increase the time spent in target  $\text{SpO}_2$  ranges in prematurely born infants (11). We have undertaken a literature review of closed loop automated oxygen control in neonates, aiming to determine their efficacy in those born prematurely or at term. Our aims were also to describe the algorithms employed and if their performance differed and whether use of closed loop automated oxygen control resulted in reductions in adverse outcomes, as well as increasing the time spent in the target oxygen saturation range.

## **METHODS**

Literature searches were completed on Google Scholar and Pubmed using the terms “closed loop” or “automat\*” and “oxygen”, “neonat\*”. Included studies were those comparing the use of closed-loop automated oxygen control systems with manual or steady-state oxygen control. The inclusion criteria were published studies that compared the use of closed-loop oxygen monitoring devices to manual or steady-state (unchanged) oxygen control with measured outcomes of time spent within target oxygen saturation range and/or number of manual adjustments to the fraction of inspired oxygen. Exclusion criteria were studies that either did not involve neonatal patients or animal studies which did not model neonatal patients’ respiratory diseases. One thousand, one hundred and forty results were identified. Abstracts were screened for relevance and duplicates removed. One hundred and forty-three studies remained, but only nineteen were studies of closed loop automated oxygen control relating to neonatal practice (Table 1).

## **RESULTS**

Eighteen studies were identified as relevant to neonatal practice: sixteen comparison, clinical studies, an observational study and an animal study. All the clinical studies were in infants of gestational ages between 23 and 30 weeks, that is there were no studies involving infants born at term. A variety of algorithms were used.

## **1. Algorithms used in closed loop automated oxygen monitoring**

### *Rule-based algorithms*

Rule-based, fuzzy algorithms operate by measuring the error and making an adjustment based on the magnitude of the error (12, 13). Fuzzy logic is used to represent the idea that a statement could vary from completely true to completely false, including being partially true or partially false, whereas non-fuzzy logic only enables a statement to be defined as 'true' or 'false'. This approach is helpful for systems which have non-linear relationships, such as the relationship between SpO<sub>2</sub> and FiO<sub>2</sub> in a neonate (14). This allows the knowledge and expertise that medical staff have to translate more easily into an algorithm (13), as staff would describe a patient's SpO<sub>2</sub> level as being slightly low or extremely low rather than 'just' low.

Rule-based algorithms determine adjustments based on the current SpO<sub>2</sub> and the trend of SpO<sub>2</sub> levels. The trend is calculated from the size of the error (how far away the SpO<sub>2</sub> is from the mid-point of the target range), its velocity and its acceleration. The trend determines what adjustment, if any, is made to the FiO<sub>2</sub> (14). For example, a closed-loop automatic control (CLAC) algorithm used in a randomised controlled trial (15) determined whether the SpO<sub>2</sub> was in the normal range, above or below (including the magnitude of the error) and whether it was increasing, decreasing or stable and then suggested five possible FiO<sub>2</sub> adjustments from -0.02 to +0.05. Additionally, the algorithm recognised when the SpO<sub>2</sub> signal was poor and excluded those readings (15). Similar rule-based algorithms have been used in preterm lambs, with significantly more time spent in target SpO<sub>2</sub> range under closed loop automated



control compared with manual control during a period of stable ventilation (post-resuscitation) (16).

#### *Proportional-integral-differential (PID) algorithms*

PID algorithms use the error, its integral and its derivative to determine the output, with multiplying coefficients (or gains) applied to each of the inputs). The error, as in rule-based algorithms, is the difference between the current SpO<sub>2</sub> and the midpoint of the target SpO<sub>2</sub> range. Some PID algorithms automatically adjust the gains over time in order to re-model the algorithm appropriately for that patient's respiratory system (17).

#### *Adaptive model algorithms*

Adaptive model control algorithms aim to model the patient's relationship between FiO<sub>2</sub> and SpO<sub>2</sub> based on the oxygen dissociation curve. The adaptive control model algorithm uses this curve to determine how much the FiO<sub>2</sub> needs to be adjusted to cause the desired change in SpO<sub>2</sub> and adjusts the model at set intervals (every two to five minutes) to make it more accurate for that particular patient (18).

#### *Comparison of algorithms*

All three algorithms have been shown to increase the amount of time spent in the target saturation range (10) and decrease the number of manual interventions required (10). The adaptive model control algorithm has been shown to result in the largest increase in time spent within the target saturation range compared to manual control (60% compared to manual control, whereas the PID algorithm achieved 52% and the rule-based algorithm achieved 56% when compared to manual control) (10).

Additionally, the adaptive model algorithms require no setup, as the algorithm automatically adjusts to the patient's response (10).

## **2 Accuracy of closed loop automated oxygen therapy**

Sixteen single or multicentre clinical studies have been undertaken comparing closed loop automated oxygen control to manual oxygen control in neonates to determine whether automating  $\text{FiO}_2$  control was associated with a greater target  $\text{SpO}_2$  achievement. They consistently demonstrated that closed loop automated oxygen control maintained the patient within their target  $\text{SpO}_2$  range a significantly greater proportion of the time than manual control (10, 12, 15, 16, 19-30). Closed loop automated oxygen control has been demonstrated to reduce hyperoxic episodes, in one study almost halving the frequency (9.3 to 4.7 episodes per 90 minutes) and the duration (19.3s to 10.1s) of hyperoxic episodes (21) and in another study reducing the median percentage time spent with an  $\text{SpO}_2 > 95\%$  from 41.9% to 19.3% ( $p < 0.001$ ) (28). A randomised crossover study found that an automated controller resulted in 'overshoot' (an exaggerated response to hypoxia leading to hyperoxia) more frequently than manual control, but the resulting episodes of hyperoxia were shorter than similar episodes under manual control (30).

A systematic review (11) of closed loop automated oxygen control studies found that automated control of  $\text{FiO}_2$  resulted in a significantly higher time spent in the target saturation range (mean difference (MD) 12.8%, 95% CI 6.5-19.2%). It also found that automated control resulted in significantly reduced periods of hyperoxia (MD: -8.8%; 95% CI: -15 to -2.7%), severe hypoxia (MD: -0.9%; 95% CI: -1.5 to -0.4%) and the number of hypoxic events (MD: -5.6; 95% CI: -9.1 to -2.1%) (11). One

study, however, found that automated control was associated with an increase in the number of episodes of hypoxia, although there was no increase in the episodes of extreme hypoxia ( $\text{SpO}_2 < 80\%$  or  $< 75\%$ ) (23).

#### *Applying automated $\text{O}_2$ under different conditions*

Automated systems have been shown to increase the percentage of time spent in the target oxygen saturation range for patients at a range of postnatal ages, despite oxygen stability varying with postnatal age. As before, none of these studies have included term born infants. They have also been shown to be effective for infants who were intubated and ventilated (10, 20, 22-24), as well as those on non-invasive respiratory support (12, 19, 21, 25, 27, 30). In a randomised trial of a closed loop automated oxygen control system for infants on non-invasive respiratory support, infants on automated control spent a significantly higher proportion of their time within their target saturation range compared to those on manual control (58% versus 33.7% respectively) (12). The automated system was also associated with a reduction in the frequency and duration of episodes of hyperoxia (12).

#### *Comparison with optimum manual control*

Automated oxygen control has been compared with routine manual control as described above, but also with optimum manual control that is one to one dedicated attention to  $\text{FiO}_2$  adjustment. Automated control was superior to one-to-one attention to oxygen adjustment in maintaining  $\text{SpO}_2$  within the target range (81% of the time versus 69%) (17). A study of automated control in preterm infants on non-invasive respiratory support found that it provided a significantly increased percentage of time within the target range when compared to 'best' manual control (78% versus

55% in the target SpO<sub>2</sub> range,  $p=0.0001$ ) (27). In that study, prolonged periods outside of the target saturation range were almost eliminated by automated control with no 60-second episodes with SpO<sub>2</sub> levels >96% or <85% and fewer than two episodes lasting 30 seconds (27). This suggests that even with the best possible staffing ratios, automated oxygen control will still confer a benefit to patients.

#### *Comparison with different oxygen saturation ranges*

A crossover trial of infants with closed loop automated SpO<sub>2</sub> control compared two target saturation ranges: 87-93% and 90-93% (31). The tighter target range narrowed the distribution of SpO<sub>2</sub> values, but did not result in any significant reduction in the frequency or duration of episodes of desaturation, increase the time spent within the target SpO<sub>2</sub> range, or decrease the time spent at extremes of SpO<sub>2</sub> (31). A further study compared two other target SpO<sub>2</sub> ranges: 89-93% and 91-95% (25). The automated oxygen control system significantly increased the time spent within the target saturation range with both target SpO<sub>2</sub> ranges compared to manual control (25). The effect was larger in those with the lower target SpO<sub>2</sub> range group (a difference of 8% compared to 4% more time in the target saturation range), although this appeared to be due to less effective manual control in that group (25).

#### *Supplementary oxygen weaning*

A crossover study found that with automated weaning, infants spent significantly more time with a FiO<sub>2</sub> <0.25 than with manual control (23). A further crossover study of automated FiO<sub>2</sub> control versus manual control in infants on nasal cannula oxygen found that, in addition to automated control resulting in a significantly greater proportion of time with saturations within the target range, the automated

control also tended to result in a FiO<sub>2</sub> of 21%, whereas manual control resulted in an FiO<sub>2</sub> of 30% (30). Those results suggest automated control might identify which patients no longer require supplementary oxygen more quickly. The results of a systematic review of automated control, however, did not demonstrate a significant difference in FiO<sub>2</sub> exposure or time spent below the target SpO<sub>2</sub> range (11). More evidence is required to determine whether automated control will aid weaning and in which groups of patients.

#### *Other outcomes*

No data were reported on adverse effects of closed-loop oxygen control, although the majority of the included studies provided the option for manual overriding of the automated systems in the case of an unsafe alteration to the FiO<sub>2</sub>. No data were found on the long-term clinical outcomes for infants with closed-loop oxygen control as compared to those with conventional manual control. A large, multicentre randomised controlled trial including long-term outcomes is underway and expects to complete in 2022 (32).

### **3. Decrease in staff workload**

Multiple studies have found that automated oxygen control systems lead to a decrease in the number of manual interventions made by staff (10, 12, 15, 20, 21, 23-25, 27, 29). Studies have reported that clinical staff needed to adjust the FiO<sub>2</sub> less than once per hour when there was automated control (10, 12, 21, 24, 27, 29), compared to as many as 29 times per hour during manual control (20). A randomised study of one automated FiO<sub>2</sub> system found that the automated system

made 7,540 FiO<sub>2</sub> adjustments in a 12-hour period, whereas staff made 80 adjustments over the same time period in the manual control group (12). Furthermore, the staff did not make any manual adjustments to the automated control group's FiO<sub>2</sub> (12), suggesting that they thought the automated adjustments were all deemed appropriate.

In a crossover study, staff adjusted the FiO<sub>2</sub> 2.3 times/hour during manual control, whereas the automated system adjusted the FiO<sub>2</sub> 64 times/hour (27). The frequency of adjustments may contribute to the overall increase in target saturation achievement with automated oxygen control (11). No evidence, however, was available to suggest that the reduction in manual interventions resulted in neonatal practitioners having a reduced workload and hence able to use that time for other care activities.

#### **4. Critical analysis of included studies**

Similarly, to Mitra's systematic review, we found a high risk of bias. This was mostly due to the fact that, although the majority of studies were randomised crossover studies, few details of the randomisation process and sequence generation were described. One study was observational, therefore, introducing a risk of confounding factors (such as gestational age or disease state) although the authors did report that the demographics of the two sets of patients were not statistically significant (28).

## DISCUSSION

This review highlights that automated oxygen control does increase the percentage of oxygen saturation levels within the target range. It should be noted, however, that even with automated systems the time infants spend in the target span of SpO<sub>2</sub> is still far from satisfactory and that there is much room for improvement. It also reduces the number of manual modifications made to the FiO<sub>2</sub>, but whether this reduces the nursing workload enabling improvements in clinical outcome by nurses being able to concentrate on other tasks has not been documented. Furthermore, whether automated oxygen control reduces long term complications has not been investigated.

There are, however, some concerns regarding automated oxygen control. Body movement is known to be a common cause of disruption of SpO<sub>2</sub> signal, but it can also be associated with hypoxaemia. Some automated systems are designed not to adjust FiO<sub>2</sub> if the SpO<sub>2</sub> signal is of a poor quality. Therefore, there could be a delay in reaction to hypoxaemia until the signal quality is restored (33). Episodes of desaturation are commonly used as a potential warning sign of infection or respiratory deterioration (34). An automated FiO<sub>2</sub> system could potentially ‘mask’ this important clinical sign by maintaining the SpO<sub>2</sub> within the target range. Therefore, some systems have additional alarm features to signal a significant increase in FiO<sub>2</sub> even if the SpO<sub>2</sub> remains within the target range. For example, one system activates an alarm if there is an increase in FiO<sub>2</sub> of  $\geq 0.3$  from the basal level (23).

In conclusion, automated oxygen control does reduce the time spent in which the oxygen saturation level is outside the target range. It is, however, important to determine whether this improves long term complications related to oxygen toxicity and if use of such systems reduces nurse workload which translates into improved clinical outcomes.



## **ACKNOWLEDGEMENTS**

**Funding source:** The research was supported by the National Institute for Health Research (NIHR) Biomedical Research Centre based at Guy's and St Thomas' NHS Foundation Trust and King's College London. The views expressed are those of the author(s) and not necessarily those of the NHS, the NIHR or the Department of Health.

**Conflict of interest:** AG has held grants from various manufacturers (Abbot Laboratories, MedImmune) and ventilator manufacturers (SLE). Professor Greenough has received honoraria for giving lectures and advising various manufacturers (Abbot Laboratories, MedImmune) and ventilator manufacturers (SLE). Professor Greenough is currently receiving a non-conditional educational grant from SLE.

## **LIST OF ABBREVIATIONS**

CLAC	Closed loop automatic control
CPAP	Continuous positive airway pressure
FiO <sub>2</sub>	Fractional concentration of inspired oxygen
MD	Mean difference
NCPAP	Nasal continuous positive airway pressure
PaO <sub>2</sub>	Partial pressure of oxygen in arterial blood
PID	Proportional integral differential
RDS	Respiratory distress syndrome
ROS	Reactive oxide species
SpO <sub>2</sub>	Oxygen saturation measured by pulse oximetry

## REFERENCES

1. Kayton A, Timoney P, Vargo L, Perez JA. A Review of oxygen physiology and appropriate management of oxygen levels in premature neonates. *Adv Neonatal Care* 2018; 18:98–104.
2. Perrone S, Bracciali C, Di Virgilio N, Buonocore G. Oxygen Use in Neonatal Care: A two-edged sword. *Front Pediatr* 2017; 4:1–7.
3. The BOOST-II Australia and United Kingdom Collaborative Groups. Outcomes of Two Trials of Oxygen-Saturation Targets in Preterm Infants. *N Engl J Med* 2016; 374:749–60.
4. Saugstad OD, Aune D. Optimal oxygenation of extremely low birth weight infants: A meta-analysis and systematic review of the oxygen saturation target studies. *Neonatology* 2013; 105:55–63.
5. Di Fiore JM, Bloom JN, Orge F, Schutt A, Schluchter M, Cheruvu VK, et al. A higher incidence of intermittent hypoxemic episodes is associated with severe retinopathy of prematurity. *J Pediatr* 2010; 157:69–73.
6. Hagadorn JI, Furey AM, Nghiem T-H, Schmid CH, Phelps DL, Pillers D-AM, et al. Achieved versus intended pulse oximeter saturation in infants born less than 28 weeks' gestation: The AVIOx study. *Pediatrics* 2006; 118:1574–82.
7. Ford SP, Leick-Rude MK, Meinert KA, Anderson B, Sheehan MB, Haney BM, et al. Overcoming Barriers to Oxygen Saturation Targeting. *Pediatrics* 2006; 118:S177–86.
8. Sink DW, Hope SAE, Hagadorn JI. Nurse: Patient ratio and achievement of oxygen saturation goals in premature infants. *Arch Dis Child Fetal Neonatal Ed* 2011; 96:96–101.
9. Clucas L, Doyle LW, Dawson J, Donath S, Davis PG. Compliance with alarm limits for pulse oximetry in very preterm infants. *Pediatrics* 2007; 119:1056–60.

10. Morozoff E, Smyth J, Saif M. Applying computer models to realize closed-loop neonatal oxygen therapy. *Anesth Analg* 2017; 124:95–103.
11. Mitra S, Singh B, El-Naggar W, McMillan DD. Automated versus manual control of inspired oxygen to target oxygen saturation in preterm infants: a systematic review and meta-analysis. *J Perinatol* 2018; 38:351–60.
12. Zapata J, Gomez JJ, Araque Campo R, Matiz Rubio A, Sola A. A randomised controlled trial of an automated oxygen delivery algorithm for preterm neonates receiving supplemental oxygen without mechanical ventilation. *Acta Paediatr* 2014; 103:928–33.
13. López JA, Araque Campo R, Matiz Rubio A. Automixer: equipment for the reduction of risks associated with inadequate oxygen supply. *Ing e Investig* 2014; 34:60–5.
14. Fathabadi OS, Gale TJ, Olivier JC, Dargaville PA. Automated control of inspired oxygen for preterm infants: What we have and what we need. *Biomed Signal Process Control* 2016; 28:9–18.
15. Hallenberger A, Poets CF, Horn W, Seyfang A, Urschitz MS. Closed-Loop Automatic Oxygen Control (CLAC) in Preterm Infants: A randomized controlled trial. *Pediatrics* 2014; 133:e379–85.
16. Hutten MC, Goos TG, Ophelders D, Nikiforou M, Kuypers E, Willems M, et al. Fully automated predictive intelligent control of oxygenation (PRICO) in resuscitation and ventilation of preterm lambs. *Pediatr Res* 2015; 78:657–63.
17. Bhutani VK, Delivoria-Papadopoulos M. Diagnostic and therapeutic methods - adaptive control of inspired oxygen delivery to the neonate. *Pediatr Pulmonol* 1992; 14:110–7.
18. Morozoff EP, Smyth JA. Evaluation of three automatic oxygen therapy control algorithms on ventilated low birth weight neonates. *Conf Proc IEEE Eng Med Biol Soc*

- 2009; 2009:3079–82.
19. Beddis IR, Collins P, Levy NM, Godfrey S, Silverman M. New technique for servo-control of arterial oxygen tension in preterm infants. *Arch Dis Child* 1979; 54:278–80.
  20. Claire N, Gerhardt T, Everett R, Musante G, Herrera C, Bancalari E. Closed-loop controlled inspired oxygen concentration for mechanically ventilated very low birth weight infants with frequent episodes of hypoxemia. *Pediatrics* 2001; 107:1120–4.
  21. Urschitz MS, Horn W, Seyfang A, Hallenberger A, Herberts T, Miksch S, et al. Automatic control of the inspired oxygen fraction in preterm infants. *Am J Respir Crit Care Med* 2004; 170:1095–100.
  22. Claire N, D’Ugard C, Bancalari E. Automated adjustment of inspired oxygen in preterm infants with frequent fluctuations in oxygenation: A pilot clinical trial. *J Pediatr* 2009; 155:640-645.e2.
  23. Claire N, Bancalari E, D’Ugard C, Nelin L, Stein M, Ramanathan R, et al. Multicenter crossover study of automated control of inspired oxygen in ventilated preterm infants. *Pediatrics* 2011; 127:e76-83.
  24. Lal M, Tin W, Sinha S. Automated control of inspired oxygen in ventilated preterm infants: crossover physiological study. *Acta Paediatr* 2015; 104:1084-9.
  25. Van Kaam AH, Hummler HD, Wilinska M, Swietlinski J, Lal MK, te Pas AB, et al. Automated versus manual oxygen control with different saturation targets and modes of respiratory support in preterm infants. *J Pediatr* 2015; 167:545-550.e2.
  26. Waitz M, Schmid MB, Fuchs H, Mendler MR, Dreyhaupt J, Hummler HD. Effects of automated adjustment of the inspired oxygen on fluctuations of arterial and regional cerebral tissue oxygenation in preterm infants with frequent desaturations. *J Pediatr* 2015; 166:240-4.e1.
  27. Plottier GK, Wheeler KI, Ali SKMM, Fathabadi OS, Jayakar R, Gale TJ, et al. Clinical

- evaluation of a novel adaptive algorithm for automated control of oxygen therapy in preterm infants on non-invasive respiratory support. *Arch Dis Child Fetal Neonatal Ed* 2017; 102:F37–43.
28. Van Zanten HA, Kuypers KLAM, Stenson BJ, Bachman TE, Pauws SC, Te Pas AB. The effect of implementing an automated oxygen control on oxygen saturation in preterm infants. *Arch Dis Child Fetal Neonatal Ed* 2017; 102:F395–9.
  29. Gajdos M, Waitz M, Mendler MR, Braun W, Hummler H. Effects of a new device for automated closed loop control of inspired oxygen concentration on fluctuations of arterial and different regional organ tissue oxygen saturations in preterm infants. *Arch Dis Child Fetal Neonatal Ed* 2019; 104:F360-5.
  30. Reynolds PR, Miller TL, Volakis LI, Holland N, Dungan GC, Roehr CC, et al. Randomised cross-over study of automated oxygen control for preterm infants receiving nasal high flow. *Arch Dis Child Fetal Neonatal Ed* 2019; 104:F366-71.
  31. Wilinska M, Bachman T, Swietlinski J, Kostro M, Twardoch-Drozdz M. Automated FiO<sub>2</sub>-SpO<sub>2</sub> control system in neonates requiring respiratory support: a comparison of a standard to a narrow SpO<sub>2</sub> control range. *BMC Pediatr* 2014; 14:130.
  32. Poets CF, Franz AR. Achieving stability in oxygenation: servo-controlled oxygen environment as a first step to fully automated oxygen control? *J Pediatr* 2018; 200:8–9.
  33. Claure N. Automated regulation of inspired oxygen in preterm infants: oxygenation stability and clinician workload. *Anesth Analg* 2007; 105:S37-41.
  34. Simonsen KA, Anderson-Berry AL, Delair SF, Dele Davies H. Early-onset neonatal sepsis. *Clin Microbiol Rev* 2014; 27:21–47.

**Table 1: Included studies**

Author (Year)	Type of study	Population	N	Algorithm	% time spent in target SpO <sub>2</sub> range		Number of manual adjustments		Results
					Manual	Automated	Manual	Automated	
Morozoff (2017) (10)	Crossover study	Low birthweight, ventilated, preterm	7	State machine	35**	56**	3.5/study period *	0/study period *	<ul style="list-style-type: none"> <li>Increased time within target range with all three algorithms tested compared to manual control</li> <li>All algorithms significantly reduced time spent in hyperoxaemia</li> <li>Algorithms did not reduce hypoxia</li> </ul>
				Proportion al-integral-derivative	41*	52*	2/ study period *	0/study period *	
				Adaptive	40**	60**	2.5/ study period **	0/study period **	
Zapata (2014) (12)	Randomised controlled trial	<30 weeks gestation, <1000g birthweight, nasal cannulae, episodes of hyperoxia & hypoxia	20	Rule-based, fuzzy	33.7**	58.0**	80/120h	0/120h	<ul style="list-style-type: none"> <li>Increased time within target range with automated control</li> <li>Reduced frequency and duration of hyperoxic episodes</li> <li>Fewer episodes of SpO<sub>2</sub> 70-75% with automated control, but increased percentage of time with SpO<sub>2</sub> &lt;85%</li> </ul>
Hallenberger (2014) (15)	Multicentre randomised controlled crossover study	<37 weeks gestation using mechanical ventilation or NCPAP; FiO <sub>2</sub> ≥0.25	34	Rule-based, non-fuzzy	61.0***	72.1***	77/24h*	52/24h*	<ul style="list-style-type: none"> <li>Significant increase in mean time spent in target SpO<sub>2</sub> range shown in 3 out of 4 participating centres</li> <li>Lowest effect seen in centre with lowest SpO<sub>2</sub> target range</li> <li>Significant reduction in time spent below the target range</li> </ul>

Hutten (2015) (16)	Randomised study	Preterm lambs from newborn resuscitation	19	Rule-based	84.0*	93.2*	13.0/h	5.7/h	<ul style="list-style-type: none"> <li>Increased time within target range with automated control during steady-state ventilation</li> <li>Automated FiO<sub>2</sub> feasible to be used during resuscitation and surfactant administration</li> </ul>
Bhutani (1992) (17)	Randomised crossover study	Preterm infants previously ventilated for RDS, studied receiving supplemental O <sub>2</sub>	14	Adaptive	69	81	-	-	<ul style="list-style-type: none"> <li>Adaptive control increased time spent within the target range compared to manual control</li> <li>Minimal overshoot seen with adaptive control</li> </ul>
Morozoff (2009) (18)	Crossover study	Low birthweight, ventilated preterm infants	7	State machine	57	71	3.74/h	0.48/h	<ul style="list-style-type: none"> <li>All three algorithms tested increased time spent within target range when compared to manual control</li> <li>As the target range increased, the three algorithms converged towards equivalent performance</li> <li>Adaptive model did not require manual tuning to adjust to changes in patient physiology</li> </ul>
				Proportional-integral-derivative	57	70	3.74/h	0.45/h	
				Adaptive	57	73	3.74/h	0.23/h	
Beddis (1979) (19)	Crossover study;	Preterm, respiratory distress syndrome receiving supplemental oxygen	12	Rule-based	72.4**	87.8**	-	-	<ul style="list-style-type: none"> <li>Increased time within target range with automated control</li> <li>Reduced time with PaO<sub>2</sub> below target range with automated control</li> <li>No significant difference in time with PaO<sub>2</sub> above target range</li> <li>Shorter duration of episodes outside of target range with automated control</li> </ul>



Claire (2001) (20)	Randomised crossover study	Preterm infants, <1500g birthweight, mechanically ventilated with frequent episodes of hypoxaemia	14	Rule-based combined with differential feedback	66.3*	74.9*	29/h	n/a	<ul style="list-style-type: none"> <li>Increased percentage of time spent in normoxaemia with automated control</li> <li>No significant difference in frequency or duration of hypoxaemic or hyperoxaemic episodes</li> <li>No significant difference in mean FiO<sub>2</sub> supplied</li> </ul>
Urschitz (2004) (21)	Randomised controlled crossover study	<34 weeks gestation, NCPAP and supplemental oxygen	12	Rule-based, non-fuzzy	81.7*	90.5*	3.0/h***	0.3/h***	<ul style="list-style-type: none"> <li>11% increase in time spent within target range with automated control</li> <li>Reduced frequency and duration of hyperoxic episodes</li> <li>Reduction in SpO<sub>2</sub> fluctuations</li> </ul>
Claire (2009) (22)	Randomised crossover study	Preterm, mechanically ventilated with frequent episodes of hypoxaemia	16	Proportional-integral-derivative	42***	58***	7/4h***	34/4h***	<ul style="list-style-type: none"> <li>Increased time within target range with automated control</li> <li>Reduced time in hyperoxaemia and time spent with SpO<sub>2</sub> above the target range with automated control</li> <li>Increased frequency of episodes with SpO<sub>2</sub> &lt;88% with automated control</li> <li>Slight decrease in duration of hypoxaemic episodes with automated control</li> </ul>
Claire (2011) (23)	Multicentre randomised crossover study	Preterm infants, mechanically ventilated, SpO <sub>2</sub> instability	34	Proportional-integral-derivative	32***	40***	112/24h***	10/24h***	<ul style="list-style-type: none"> <li>Increased time within target range with automated control</li> <li>Reduced duration of hyperoxia</li> <li>Consistently lower FiO<sub>2</sub> throughout the study period with automated FiO<sub>2</sub></li> <li>Increased frequency of hypoxia</li> </ul>

Lal (2015) (24)	Randomised crossover study	Preterm, mechanically ventilated	27	Proportional-integral-derivative	59.6*		72.8*	63/12h**	0/12h**	<ul style="list-style-type: none"><li>Increased time within target range with automated control</li><li>Reduced time with SpO<sub>2</sub> &gt;98% or &lt;80%</li><li>Reduced prolonged episodes of hypoxaemia and hyperoxaemia (≥1 or ≥3 minutes)</li></ul>
Van Kaam (2015) (25)	Multicentre randomised crossover study	<33 weeks gestation, mechanically ventilated or receiving non-invasive respiratory support	80	Proportional-integral-derivative	SpO <sub>2</sub> target 89-93%	54**	62**	-	-	<ul style="list-style-type: none"><li>Mean within subject difference was significantly larger in the 89-93% target group</li><li>Reduced time in hypoxaemia and hyperoxaemia</li><li>In the 89-93% target group only: reduced time above target range, reduced severe hyperoxaemia and reduced episodes of prolonged hyperoxaemia</li></ul>
					SpO <sub>2</sub> target 91-95%	58*	62*			
Waitz (2015) (26)	Randomised crossover study	<30 weeks gestation, mechanically ventilated or CPAP/NCPAP, intermittent hypoxaemia	15	Proportional-integral-derivative	69.1**		76.3**	-	-	<ul style="list-style-type: none"><li>Increased time within target range with automated control</li><li>Reduced time in hyperoxaemia</li><li>Reduced prolonged periods with SpO<sub>2</sub> &lt;88%</li><li>No significant difference in cerebral tissue oxygen saturation between the two methods</li></ul>
Plottier (2017) (27)	Crossover study	<37 weeks gestation, non-invasive respiratory support	20	Proportional-integral-derivative	55****		78****	2.3/h	0.24/h	<ul style="list-style-type: none"><li>Increased time within target range with automated control</li><li>Reduced SpO<sub>2</sub> coefficient of variation</li><li>Reduced time spent with SpO<sub>2</sub> &lt; 80% or &gt;98%</li><li>No overshoot with automated control</li></ul>
Van Zanten (2017) (28)	Prospective observational study	<30 weeks gestation, mechanically ventilated or non-invasively	42	Proportional-integral-derivative	48.4**		62.0**	-	-	<ul style="list-style-type: none"><li>Increased time within target range with automated control</li><li>Reduced time with SpO<sub>2</sub> above target range</li><li>Increased time with SpO<sub>2</sub> between 80-</li></ul>

		ventilated							90% but no increase in time with SpO <sub>2</sub> <80%
Gajdos (2018) (29)	Randomised crossover study	<30 weeks gestation, non-invasive or invasive ventilation, SpO <sub>2</sub> instability	12	Proportional-integral-derivative	68.52**	77.83**	7.5/h***	0.5/h***	<ul style="list-style-type: none"> <li>Reduced time in hypoxaemia, including a reduction in prolonged periods of hypoxaemia &gt;60s and &gt;180s duration</li> <li>No significant difference in time spent above the target SpO<sub>2</sub> range</li> <li>No significant difference in hepatic or cerebral tissue oxygen saturation between the two methods</li> </ul>
Reynolds (2018) (30)	Two-centre randomised crossover study	Preterm, receiving high-flow oxygen	30	Used discrete target value, SpO <sub>2</sub> -FiO <sub>2</sub> relationship and dynamics of response	49****	80****	1.6/h****	0/h****	<ul style="list-style-type: none"> <li>Increased time within target range with automated control</li> <li>Reduced variation of SpO<sub>2</sub> values</li> <li>Reduced time below target range and with SpO<sub>2</sub> &lt;80%</li> <li>Reduced time with SpO<sub>2</sub> above target range</li> <li>No significant difference in frequency of episodes of hyperoxaemia</li> </ul>

\* = p < 0.05

\*\* = p < 0.01

\*\*\* - p<0.001

\*\*\*\*- p<0.0001